



April 24, 2019

KURT TOWNSON
TOWNSEND DESIGN THUASNE USA
4615 SHEPARD STREET
BAKERSFELD, CA 93313

DCN Number:19092C22100004

Manufacturer Name	Product Name	Model Number	HCPCS Code Assigned
TOWNSEND DESIGN THUASNE USA	TOWNSEND NIGHT SPLINT	24-0XX	L4396
TOWNSEND DESIGN THUASNE USA	TOWNSEND NIGHT SPLINT	24-0XX	L4397

Dear KURT TOWNSON,

The Pricing, Data Analysis, and Coding (PDAC) Contractor has reviewed the product(s) listed above and has approved the listed Healthcare Common Procedure Coding System (HCPCS) code(s) for billing the four Durable Medical Equipment Medicare Administrative Contractors (DME MACs).

The PDAC Contractor provides coding assistance to manufacturers to ensure proper coding of Durable Medical Equipment, Prosthetics, Orthotics, and Supplies (DMEPOS). The PDAC publishes coding decisions based on the coding guidelines established by the Local Coverage Determinations (LCDs) and associated Policy Articles and any related Advisory Articles established by the DME MACs. All products submitted to the PDAC for a coding verification review are examined by coders and professionals following a formal, standardized process.

Based on this review and application of DME MAC policy, the HCPCS code(s) listed below should be used when billing the DME MACs:



- L4396 STATIC OR DYNAMIC ANKLE FOOT ORTHOSIS, INCLUDING SOFT INTERFACE MATERIAL, ADJUSTABLE FOR FIT, FOR POSITIONING, MAY BE USED FOR MINIMAL AMBULATION, PREFABRICATED ITEM THAT HAS BEEN TRIMMED, BENT, MOLDED, ASSEMBLED, OR OTHERWISE CUSTOMIZED TO FIT A SPECIFIC PATIENT BY AN INDIVIDUAL WITH EXPERT
- L4397 STATIC OR DYNAMIC ANKLE FOOT ORTHOSIS, INCLUDING SOFT INTERFACE MATERIAL, ADJUSTABLE FOR FIT, FOR POSITIONING, MAY BE USED FOR MINIMAL AMBULATION, PREFABRICATED, OFF-THE-SHELF

If you disagree with this decision, you may request a reconsideration within 45 days of the letter's date and provide evidence to substantiate a reconsideration of PDAC's original coding determination. To request a reconsideration, complete the Reconsideration Request form located on the PDAC website at <https://www.dmepdac.com>. If your request for a reconsideration is made after the 45-day time frame, it will require a new application and documentation to support the request.

It is the responsibility of manufacturers and distributors to notify the PDAC immediately of any changes involving their products, as listed on the PCL on DMECS. Further information for requesting updates to the PCL can be found on the PDAC website at <https://www.dmepdac.com>. It is also the responsibility of manufacturers and distributors to assure their websites and product marketing materials accurately reflect the product reviewed by the PDAC and the coding decision assigned.

An assignment of the HCPCS code(s) to product(s) is not an approval or endorsement of the product(s) by Medicare or Palmetto GBA; nor does it imply or guarantee claim reimbursement or coverage.

If you have questions about policy, claim coverage or reimbursement, please contact the DME MAC for your jurisdiction. For other questions, contact the PDAC HCPCS Helpline at the address listed above or by telephone at (877) 735-1326. The Contact Center is open Monday through Friday from 9:30 a.m. to 5:00 p.m. EST.

Sincerely,

Pricing, Data Analysis, and Coding Contract (PDAC)
Palmetto GBA, LLC
www.dmepdac.com